
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006106
13:30 14/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:18
STERILIZING 4:28
136.4 C 223kPa 4:28
136.4 C 224kPa 4:58
136.4 C 223kPa 5:28
136.4 C 225kPa 5:58
136.0 C 222kPa 6:28
136.4 C 223kPa 6:58
136.0 C 222kPa 7:28
136.4 C 223kPa 7:58
136.4 C 223kPa 7:58
VENTING 7:58
AIR DRYING 8:48
CYCLE COMPLETE 17:03

Digital Signature #
A994D656416C05E9
