
Midmark M11 - v1.0.0
Total Cycles: 1211

Name: RITTER 1
Sterilizer ID: V2178198
09/28/2021 10:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.0 F 0.0
2:00 158.2 F 1.5
4:00 173.5 F 3.8
6:00 204.5 F 7.7
8:00 211.2 F 13.1
10:00 223.6 F 6.5
12:00 235.7 F 11.2
14:00 235.5 F 8.2
16:00 242.8 F 11.5
18:00 245.5 F 13.0
20:00 257.5 F 19.5
22:00 268.4 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.7 F 29.2
2:00 272.0 F 28.6
3:00 271.5 F 28.5
4:00 272.8 F 29.2

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.1 F 0.0
5:00 209.9 F 0.0
10:00 214.4 F 0.0
15:00 217.4 F 0.0
20:00 214.5 F 0.0
25:00 211.4 F 0.0
30:00 191.7 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 23:16
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:01:47
CYCLE COMPLETE

09/28/2021 11:04 AM
ACCEPTED: X
REJECTED:
Operator: U01