
Midmark M11 - v1.0.0
Total Cycles: 4726

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2024 08:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.8 F 0.0
2:00 72.4 F 0.1
4:00 88.7 F 0.4
6:00 120.1 F 2.5
8:00 149.4 F 5.1
10:00 172.2 F 8.3
12:00 190.4 F 11.9
14:00 202.0 F 2.7
16:00 214.8 F 6.0
18:00 224.2 F 9.4
20:00 234.1 F 13.3
22:00 234.2 F 9.1
24:00 240.9 F 11.0
26:00 243.4 F 12.2
28:00 255.1 F 18.2
30:00 265.5 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 28.8
2:00 271.7 F 28.7
3:00 272.1 F 28.7
4:00 271.8 F 28.8

Min 271.3 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.6 F 0.0
5:00 202.1 F 0.0
10:00 208.6 F 0.0
15:00 216.7 F 0.0
20:00 214.2 F 0.0
25:00 212.5 F 0.0
30:00 213.7 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:50
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:10:08
CYCLE COMPLETE

07/15/2024 09:45 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____