

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001950  
7:15 07/07/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:08  
PRESSURIZING 2:13  
STERILIZING 3:10  
136.4 C 220kPa 3:10  
136.3 C 222kPa 3:40  
136.2 C 221kPa 4:10  
136.4 C 220kPa 4:40  
136.0 C 218kPa 5:10  
136.1 C 222kPa 5:40  
136.1 C 220kPa 6:10  
136.4 C 222kPa 6:40  
136.4 C 222kPa 6:40  
VENTING 6:41  
AIR DRYING 7:28  
CYCLE COMPLETE 7:38

-----  
Digital Signature #  
5E1D006AF05CE362  
-----