
Midmark M11 - v1.0.5
Total Cycles: 2863

Name: RITTER2
Sterilizer ID: V2313228
09/11/2023 03:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.4 F 0.0
2:00 176.9 F 3.4
4:00 196.1 F 8.0
6:00 212.0 F 12.6
8:00 223.8 F 7.2
10:00 235.8 F 12.1
12:00 236.7 F 9.8
14:00 238.0 F 9.5
16:00 249.0 F 15.1
18:00 260.3 F 21.4
20:00 270.4 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.5 F 29.7
2:00 272.0 F 29.1
3:00 271.3 F 29.1
4:00 272.6 F 29.7

Min 271.2 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 210.8 F 0.0
10:00 216.0 F 0.0
15:00 214.2 F 0.0
20:00 211.0 F 0.0
25:00 208.6 F 0.0
30:00 206.0 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 20:54
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:00:41
CYCLE COMPLETE

09/11/2023 04:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____