
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002918
11:01 10/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 3:44
STERILIZING 4:44
136.4 C 222kPa 4:44
136.4 C 224kPa 5:14
136.4 C 223kPa 5:44
136.1 C 221kPa 6:14
136.4 C 223kPa 6:44
136.2 C 223kPa 7:14
136.2 C 224kPa 7:44
136.4 C 222kPa 8:14
136.4 C 222kPa 8:14
VENTING 8:14
AIR DRYING 9:11
CYCLE COMPLETE 9:14

Digital Signature #
3AB2042AFF885CF1
