
Midmark M11 - v1.0.0
Total Cycles: 4231

Name: RITTER 1
Sterilizer ID: V2178198
01/25/2024 11:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.7 F 0.0
2:00 168.2 F 1.7
4:00 167.7 F 5.1
6:00 182.4 F 8.7
8:00 202.3 F 12.8
10:00 215.4 F 5.1
12:00 225.9 F 9.1
14:00 236.9 F 13.4
16:00 236.3 F 9.9
18:00 240.4 F 9.4
20:00 246.6 F 13.5
22:00 257.6 F 19.5
24:00 267.6 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.7 F 29.0
2:00 271.6 F 28.4
3:00 272.1 F 28.9
4:00 272.4 F 28.9

Min 271.2 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 214.6 F 0.0
10:00 222.1 F 0.0
15:00 230.2 F 0.0
20:00 228.6 F 0.0
25:00 231.4 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 25:28
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:04:19
CYCLE COMPLETE

01/25/2024 12:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____