

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1860

Name: RITTER2  
Sterilizer ID: V2313228  
12/20/2022 08:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 77.4 F 0.0  
2:00 78.8 F 0.0  
4:00 150.0 F 0.0  
6:00 179.3 F 0.0  
8:00 204.8 F 0.0  
10:00 208.7 F 1.3  
12:00 203.3 F 4.8  
14:00 213.4 F 8.3  
16:00 225.9 F 12.2  
18:00 231.9 F 7.4  
20:00 242.7 F 12.4  
22:00 239.6 F 10.2  
24:00 237.0 F 8.6  
26:00 250.1 F 15.5  
28:00 261.6 F 22.0  
30:00 272.4 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.4 F 29.2  
2:00 271.5 F 28.5  
3:00 272.6 F 29.4  
4:00 271.8 F 28.6

Min 271.3 F 28.4  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.5 F 0.0  
5:00 207.4 F 0.0  
10:00 219.8 F 0.0  
15:00 229.3 F 0.0  
20:00 225.9 F 0.0  
25:00 237.3 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:29  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:08:18  
CYCLE COMPLETE

12/20/2022 10:02 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_