

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000208  
16:36 21/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:28  
PRESSURIZING 3:36  
STERILIZING 5:07  
136.0 C 229kPa 5:07  
136.4 C 233kPa 5:37  
136.5 C 232kPa 6:07  
136.4 C 233kPa 6:37  
136.3 C 233kPa 7:07  
136.3 C 232kPa 7:37  
136.3 C 232kPa 8:07  
136.3 C 231kPa 8:37  
136.3 C 231kPa 8:37  
VENTING 8:37  
AIR DRYING 9:41  
CYCLE COMPLETE 9:43  
-----

Digital Signature #  
7B59B16A83FC53B3  
-----