

-----  
Midmark M11 - v1.0.0  
Total Cycles: 983

Name: RITTER 1  
Sterilizer ID: V2178198  
07/08/2021 10:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.6 F 0.0  
2:00 183.6 F 3.8  
4:00 202.6 F 8.8  
6:00 220.3 F 8.0  
8:00 232.6 F 8.9  
10:00 237.6 F 10.3  
12:00 243.1 F 12.0  
14:00 240.8 F 10.5  
16:00 254.4 F 17.3  
18:00 266.3 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 272.3 F 28.3  
2:00 272.5 F 28.8  
3:00 271.5 F 28.0  
4:00 272.7 F 28.7  
  
Min 271.2 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.0 F 0.0  
5:00 199.6 F 0.0  
10:00 209.1 F 0.0  
15:00 204.8 F 0.0  
20:00 210.8 F 0.0  
25:00 213.4 F 0.0  
30:00 214.4 F 0.0  
Drying Complete

FILLING: 2:01  
HEATING: 19:30  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:58:36  
CYCLE COMPLETE

07/08/2021 11:10 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02