
Midmark M11 - v1.0.0
Total Cycles: 624

Name: MIDMARK
Sterilizer ID: V2178198
01/25/2021 12:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.2 F 0.0
2:00 132.2 F 0.9
4:00 137.8 F 3.1
6:00 194.1 F 7.2
8:00 199.1 F 13.0
10:00 224.6 F 6.9
12:00 231.4 F 12.4
14:00 236.8 F 9.3
16:00 237.6 F 8.7
18:00 248.4 F 14.3
20:00 261.2 F 21.5
22:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 272.6 F 28.8
2:00 271.4 F 28.4
3:00 272.1 F 28.4
4:00 272.1 F 29.0

Min 271.1 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 223.8 F 0.0
10:00 223.0 F 0.0
15:00 224.2 F 0.0
20:00 224.1 F 0.0
25:00 223.7 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 22:14
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 01:00:10
CYCLE COMPLETE

01/25/2021 01:20 PM
ACCEPTED: X
REJECTED:
Operator: U01