
Midmark M11 - v1.0.5
Total Cycles: 443

Name: RITTER2
Sterilizer ID: V2313228
10/13/2021 09:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 85.9 F 0.0
2:00 86.5 F 0.1
4:00 90.5 F 1.1
6:00 107.6 F 3.4
8:00 138.0 F 6.5
10:00 167.3 F 9.9
12:00 187.8 F 13.4
14:00 212.8 F 5.7
16:00 219.9 F 9.4
18:00 233.8 F 12.5
20:00 237.7 F 10.3
22:00 234.7 F 9.0
24:00 247.4 F 14.0
26:00 258.0 F 19.7
28:00 267.6 F 25.9
30:00 272.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.1
1:00 271.6 F 29.2
2:00 271.8 F 29.1
3:00 271.9 F 29.0
4:00 272.5 F 29.2

Min 271.4 F 29.0
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 227.8 F 0.0
10:00 206.7 F 0.0
15:00 205.7 F 0.0
20:00 218.0 F 0.0
25:00 202.3 F 0.0
30:00 191.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 30:03
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:08:16
CYCLE COMPLETE

10/13/2021 10:36 AM
ACCEPTED: X
REJECTED:
Operator: U01