
Midmark M11 - v1.0.5
Total Cycles: 1747

Name: RITTER2
Sterilizer ID: V2313228
11/17/2022 06:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.1 F 0.0
2:00 182.2 F 0.6
4:00 200.7 F 5.1
6:00 219.0 F 10.4
8:00 225.5 F 7.0
10:00 239.5 F 12.6
12:00 240.4 F 11.3
14:00 240.7 F 11.0
16:00 255.0 F 18.2
18:00 267.2 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.5 F 29.2
2:00 271.6 F 29.0
3:00 272.5 F 29.3
4:00 271.5 F 28.6

Min 271.5 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 210.9 F 0.0
10:00 223.6 F 0.0
15:00 230.1 F 0.0
20:00 229.0 F 0.0
25:00 227.6 F 0.0
30:00 226.4 F 0.0

Drying Complete

FILLING: 1:39
HEATING: 19:22
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:18
CYCLE COMPLETE

11/17/2022 06:59 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____