

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 004356  
9:45 14/09/2022

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:35  
STERILIZING 3:28  
136.0 C 220kPa 3:28  
136.4 C 225kPa 3:58  
136.4 C 224kPa 4:28  
136.2 C 223kPa 4:58  
136.4 C 224kPa 5:28  
136.4 C 223kPa 5:58  
136.4 C 222kPa 6:28  
136.3 C 222kPa 6:58  
136.3 C 222kPa 6:58  
VENTING 6:58  
AIR DRYING 8:01  
CYCLE COMPLETE 12:15  
-----

Digital Signature #  
17266A396934EA53  
-----