
Midmark M11 - v1.0.0
Total Cycles: 4609

Name: RITTER 1
Sterilizer ID: V2178198
06/05/2024 10:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.2 F 0.0
2:00 126.9 F 1.2
4:00 140.3 F 3.7
6:00 162.7 F 7.2
8:00 220.8 F 11.8
10:00 215.0 F 4.2
12:00 227.3 F 8.2
14:00 238.8 F 12.6
16:00 233.8 F 8.6
18:00 243.8 F 12.1
20:00 243.2 F 11.8
22:00 255.3 F 18.1
24:00 266.3 F 24.7
26:00 272.8 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 272.8 F 29.2
2:00 271.3 F 28.4
3:00 272.8 F 29.1
4:00 271.3 F 28.5

Min 271.3 F 28.4
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 217.7 F 0.0
10:00 196.7 F 0.0
15:00 175.4 F 0.0
20:00 173.8 F 0.0
25:00 176.7 F 0.0
30:00 176.2 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 26:15
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:04:31
CYCLE COMPLETE

06/05/2024 11:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____