
Midmark M11 - v1.0.5
Total Cycles: 4155

Name: RITTER2
Sterilizer ID: V2313228
11/22/2024 02:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.9 F 0.0
2:00 128.1 F 0.0
4:00 160.6 F 0.4
6:00 189.3 F 3.5
8:00 205.4 F 7.3
10:00 218.2 F 11.2
12:00 224.9 F 5.8
14:00 233.4 F 10.6
16:00 235.4 F 8.3
18:00 243.4 F 13.4
20:00 243.2 F 12.5
22:00 254.7 F 18.6
24:00 265.4 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.5 F 30.1
2:00 271.9 F 29.4
3:00 271.5 F 29.4
4:00 272.6 F 30.1

Min 271.3 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 199.8 F 0.0
10:00 221.3 F 0.0
15:00 218.5 F 0.0
20:00 217.8 F 0.0
25:00 224.6 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:54
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:04:38
CYCLE COMPLETE

11/22/2024 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____