
Midmark M11 - v1.0.0
Total Cycles: 2660

Name: RITTER 1
Sterilizer ID: V2178198
12/08/2022 12:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.2 F 0.0
2:00 182.7 F 3.9
4:00 193.3 F 8.4
6:00 213.6 F 13.5
8:00 224.7 F 8.4
10:00 237.5 F 11.8
12:00 239.4 F 11.2
14:00 237.9 F 9.4
16:00 251.9 F 16.3
18:00 263.9 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.5
1:00 272.6 F 29.3
2:00 271.9 F 28.5
3:00 272.7 F 29.3
4:00 272.1 F 28.6

Min 271.5 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.3 F 0.0
5:00 231.2 F 0.0
10:00 236.9 F 0.0
15:00 231.9 F 0.0
20:00 223.5 F 0.0
25:00 220.1 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 19:51
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:08
CYCLE COMPLETE

12/08/2022 01:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____