
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004823
9:50 07/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:55
PRESSURIZING 2:58
STERILIZING 4:06
136.4 C 223kPa 4:06
136.2 C 222kPa 4:36
136.4 C 223kPa 5:06
136.4 C 223kPa 5:36
136.4 C 222kPa 6:06
136.4 C 220kPa 6:36
136.4 C 223kPa 7:06
136.4 C 222kPa 7:36
136.4 C 222kPa 7:36
VENTING 7:36
AIR DRYING 8:27
CYCLE COMPLETE 10:30

Digital Signature #
EAE87BC4DCF835B7
