

-----  
Midmark M11 - v1.0.0  
Total Cycles: 987

Name: RITTER 1  
Sterilizer ID: V2178198  
07/08/2021 02:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.3 F 0.0  
2:00 192.4 F 3.7  
4:00 215.3 F 9.3  
6:00 227.4 F 7.3  
8:00 238.5 F 11.1  
10:00 236.6 F 8.9  
12:00 238.5 F 9.3  
14:00 250.8 F 15.4  
16:00 263.8 F 23.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 271.7 F 28.0  
2:00 272.3 F 28.4  
3:00 272.6 F 28.8  
4:00 271.6 F 28.2  
  
Min 271.4 F 27.9  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.3 F 0.0  
5:00 203.9 F 0.0  
10:00 217.7 F 0.0  
15:00 224.1 F 0.0  
20:00 217.1 F 0.0  
Dry Cycle Aborted

FILLING: 3:21  
HEATING: 17:49  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 24:08  
TOTAL CYCLE: 00:52:22

07/08/2021 03:14 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02