
Midmark M11 - v1.0.5
Total Cycles: 4335

Name: RITTER2
Sterilizer ID: V2313228
02/05/2025 02:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.5 F 0.0
2:00 92.2 F 0.2
4:00 99.7 F 1.9
6:00 123.8 F 4.6
8:00 159.6 F 8.0
10:00 179.0 F 11.9
12:00 204.9 F 3.7
14:00 214.0 F 7.3
16:00 228.6 F 11.2
18:00 228.6 F 7.0
20:00 240.6 F 12.3
22:00 239.1 F 10.4
24:00 250.8 F 16.2
26:00 260.7 F 22.2
28:00 271.3 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.7
1:00 272.1 F 29.6
2:00 272.2 F 29.6
3:00 272.4 F 29.7
4:00 272.5 F 29.9

Min 271.8 F 29.6
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.5 F 0.0
5:00 216.0 F 0.0
10:00 185.3 F 0.0
15:00 181.6 F 0.0
20:00 173.6 F 0.0
25:00 171.0 F 0.0
30:00 169.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:28
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:08:02
CYCLE COMPLETE

02/05/2025 03:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____