
Midmark M11 - v1.0.5
Total Cycles: 4179

Name: RITTER2
Sterilizer ID: V2313228
12/03/2024 04:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.1 F 0.0
2:00 153.4 F 0.2
4:00 175.4 F 1.6
6:00 197.7 F 5.3
8:00 210.5 F 9.4
10:00 224.6 F 13.6
12:00 230.2 F 8.7
14:00 241.2 F 13.5
16:00 241.2 F 11.6
18:00 240.1 F 10.6
20:00 253.1 F 17.1
22:00 264.3 F 23.8
24:00 271.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.6 F 29.5
2:00 272.2 F 28.9
3:00 271.3 F 28.8
4:00 272.5 F 29.6

Min 271.2 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.3 F 0.0
5:00 215.6 F 0.0
10:00 221.2 F 0.0
15:00 214.6 F 0.0
20:00 205.5 F 0.0
25:00 190.6 F 0.0
30:00 187.1 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 24:04
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:03:04
CYCLE COMPLETE

12/03/2024 06:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____