
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003981
9:10 11/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:04
STERILIZING 3:04
136.4 C 220kPa 3:04
136.2 C 223kPa 3:34
136.2 C 223kPa 4:04
136.4 C 219kPa 4:34
136.3 C 222kPa 5:04
136.4 C 220kPa 5:34
136.3 C 218kPa 6:04
136.2 C 222kPa 6:34
136.2 C 222kPa 6:34
VENTING 6:34
AIR DRYING 7:20
CYCLE COMPLETE 7:22

Digital Signature #
293A2DB004AEA1ED
