
Midmark M11 - v1.0.0
Total Cycles: 5253

Name: RITTER 1
Sterilizer ID: V2178198
01/17/2025 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.0 F 0.0
2:00 182.5 F 1.3
4:00 189.3 F 5.0
6:00 198.6 F 9.2
8:00 215.6 F 10.1
10:00 222.3 F 7.1
12:00 234.6 F 11.9
14:00 235.1 F 8.9
16:00 241.0 F 10.5
18:00 246.3 F 13.3
20:00 258.2 F 19.7
22:00 268.9 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 29.0
2:00 271.9 F 28.4
3:00 271.7 F 28.5
4:00 272.6 F 29.0

Min 271.4 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 205.8 F 0.0
10:00 211.8 F 0.0
15:00 224.5 F 0.0
20:00 229.6 F 0.0
25:00 231.4 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 23:11
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:01:51
CYCLE COMPLETE

01/17/2025 12:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____