

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4399

Name: RITTER 1  
Sterilizer ID: V2178198  
03/22/2024 02:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.4 F 0.0  
2:00 144.4 F 2.3  
4:00 159.1 F 5.5  
6:00 182.6 F 9.4  
8:00 201.7 F 10.5  
10:00 214.9 F 6.5  
12:00 228.6 F 11.1  
14:00 227.8 F 6.9  
16:00 241.2 F 12.5  
18:00 239.7 F 10.4  
20:00 252.0 F 16.5  
22:00 262.9 F 22.7  
24:00 272.8 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.5 F 29.2  
2:00 271.8 F 28.5  
3:00 271.8 F 29.0  
4:00 272.5 F 29.1

Min 271.2 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.9 F 0.0  
5:00 207.7 F 0.0  
10:00 212.9 F 0.0  
15:00 214.0 F 0.0  
20:00 220.1 F 0.0  
25:00 221.6 F 0.0  
30:00 221.1 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 24:22  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:03:02  
CYCLE COMPLETE

03/22/2024 03:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_