
Midmark M11 - v1.0.0
Total Cycles: 4941

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2024 10:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.7 F 0.0
2:00 106.0 F 0.0
4:00 113.9 F 0.1
6:00 137.3 F 0.2
8:00 175.7 F 0.4
10:00 194.0 F 2.9
12:00 212.3 F 6.1
14:00 225.5 F 9.8
16:00 232.6 F 11.4
18:00 232.4 F 9.3
20:00 241.0 F 11.0
22:00 242.2 F 11.8
24:00 239.5 F 10.1
26:00 252.0 F 16.2
28:00 262.6 F 22.3
30:00 272.2 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.2 F 28.8
2:00 272.4 F 28.8
3:00 272.0 F 28.4
4:00 271.5 F 28.1

Min 271.3 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 218.8 F 0.0
10:00 212.4 F 0.0
15:00 233.4 F 0.0
20:00 227.8 F 0.0
25:00 231.1 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 30:36
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:09:08
CYCLE COMPLETE

09/26/2024 11:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____