
Midmark M11 - v1.0.5
Total Cycles: 3308

Name: RITTER2
Sterilizer ID: V2313228
01/19/2024 10:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.2 F 0.0
2:00 144.0 F 0.7
4:00 149.1 F 3.0
6:00 171.4 F 6.6
8:00 194.9 F 10.8
10:00 213.0 F 4.6
12:00 227.9 F 8.0
14:00 237.6 F 12.7
16:00 237.5 F 9.8
18:00 237.8 F 9.5
20:00 248.2 F 14.6
22:00 259.7 F 20.9
24:00 270.1 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.7 F 29.3
2:00 271.3 F 28.5
3:00 272.4 F 29.4
4:00 271.9 F 28.9

Min 271.1 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.1 F 0.0
5:00 226.8 F 0.0
10:00 223.1 F 0.0
15:00 199.0 F 0.0
20:00 196.7 F 0.0
25:00 172.5 F 0.0
30:00 170.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:56
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:03:30
CYCLE COMPLETE

01/19/2024 11:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____