
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002601
8:24 25/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 3:48
STERILIZING 4:59
136.0 C 220kPa 4:59
136.0 C 222kPa 5:29
136.3 C 223kPa 5:59
136.2 C 221kPa 6:29
136.4 C 224kPa 6:59
136.1 C 223kPa 7:29
136.3 C 224kPa 7:59
136.2 C 222kPa 8:29
136.2 C 222kPa 8:29
VENTING 8:29
AIR DRYING 9:19
CYCLE COMPLETE 10:40

Digital Signature #
BB1D127CA3202DD0
