
Midmark M11 - v1.0.5
Total Cycles: 1744

Name: RITTER2
Sterilizer ID: V2313228
11/17/2022 02:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.3 F 0.0
2:00 169.7 F 0.3
4:00 190.5 F 2.7
6:00 215.2 F 7.7
8:00 228.5 F 13.1
10:00 235.7 F 9.8
12:00 235.6 F 8.8
14:00 241.2 F 11.0
16:00 248.3 F 14.6
18:00 261.2 F 21.9
20:00 272.8 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.6 F 29.3
2:00 271.3 F 28.7
3:00 272.5 F 29.4
4:00 271.8 F 28.7

Min 271.2 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.9 F 0.0
5:00 224.0 F 0.0
10:00 232.8 F 0.0
15:00 230.5 F 0.0
20:00 228.2 F 0.0
25:00 227.4 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 20:23
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

11/17/2022 03:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____