
Midmark M11 - v1.0.0
Total Cycles: 3886

Name: RITTER 1
Sterilizer ID: V2178198
10/11/2023 07:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 71.1 F 0.0
4:00 79.6 F 0.0
6:00 114.1 F 0.1
8:00 154.7 F 1.3
10:00 209.5 F 4.6
12:00 201.6 F 8.2
14:00 214.9 F 11.5
16:00 218.3 F 4.1
18:00 230.7 F 8.2
20:00 237.6 F 12.4
22:00 235.0 F 8.7
24:00 241.8 F 11.4
26:00 244.6 F 12.4
28:00 256.5 F 18.8
30:00 267.5 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.4 F 28.5
2:00 272.0 F 28.9
3:00 271.7 F 28.2
4:00 272.4 F 29.0

Min 271.1 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.7 F 0.0
5:00 238.8 F 0.0
10:00 224.3 F 0.0
15:00 228.9 F 0.0
20:00 228.3 F 0.0
25:00 228.2 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:25
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:09:35
CYCLE COMPLETE

10/11/2023 09:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____