
Midmark M11 - v1.0.5
Total Cycles: 2752

Name: RITTER2
Sterilizer ID: V2313228
08/10/2023 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 213.3 F 0.0
2:00 199.1 F 4.2
4:00 218.1 F 10.1
6:00 229.0 F 6.7
8:00 238.7 F 12.4
10:00 239.4 F 10.9
12:00 238.5 F 10.2
14:00 253.1 F 17.4
16:00 265.2 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 272.7 F 29.6
2:00 271.3 F 29.0
3:00 272.5 F 29.4
4:00 271.3 F 29.1

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 230.4 F 0.0
10:00 227.3 F 0.0
15:00 225.0 F 0.0
20:00 221.1 F 0.0
25:00 220.7 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 2:40
HEATING: 17:37
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:31
CYCLE COMPLETE

08/10/2023 01:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____