
Midmark M11 - v1.0.0
Total Cycles: 3098

Name: RITTER 1
Sterilizer ID: V2178198
04/12/2023 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.2 F 0.0
2:00 174.1 F 0.5
4:00 182.9 F 4.0
6:00 203.4 F 8.9
8:00 219.9 F 10.1
10:00 230.0 F 9.2
12:00 237.3 F 9.7
14:00 243.1 F 12.8
16:00 242.3 F 11.7
18:00 255.7 F 18.6
20:00 267.5 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.6 F 29.2
2:00 271.6 F 28.5
3:00 272.6 F 29.3
4:00 272.0 F 28.7

Min 271.6 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.9 F 0.0
5:00 209.6 F 0.0
10:00 240.6 F 0.0
15:00 227.8 F 0.0
20:00 219.8 F 0.0
25:00 220.2 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 21:21
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

04/12/2023 10:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____