
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000940
16:50 17/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:49
PRESSURIZING 3:52
STERILIZING 5:06
136.0 C 227kPa 5:06
136.3 C 232kPa 5:36
136.4 C 233kPa 6:06
136.4 C 233kPa 6:36
136.3 C 232kPa 7:06
136.4 C 231kPa 7:36
136.3 C 231kPa 8:06
136.4 C 233kPa 8:36
136.4 C 233kPa 8:36
VENTING 8:36
AIR DRYING 9:29
CYCLE COMPLETE 10:53

Digital Signature #
8E83B36F34E18B13
