
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003462
10:10 18/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:24
PRESSURIZING 1:27
STERILIZING 2:29
136.5 C 223kPa 2:29
136.3 C 225kPa 2:59
136.2 C 223kPa 3:29
136.1 C 218kPa 3:59
136.5 C 225kPa 4:29
136.4 C 224kPa 4:59
136.4 C 224kPa 5:29
136.2 C 221kPa 5:59
136.2 C 221kPa 5:59
VENTING 5:59
AIR DRYING 6:59
CYCLE COMPLETE 12:30

Digital Signature #
AF3AE07842BCFA71
