
Midmark M11 - v1.0.5
Total Cycles: 1074

Name: RITTER2
Sterilizer ID: V2313228
05/05/2022 06:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.4 F 0.0
2:00 74.6 F 0.0
4:00 84.9 F 0.0
6:00 102.8 F 0.1
8:00 153.3 F 0.3
10:00 180.8 F 0.9
12:00 200.1 F 4.1
14:00 212.5 F 7.5
16:00 222.1 F 11.4
18:00 225.6 F 6.5
20:00 237.4 F 11.4
22:00 235.5 F 9.0
24:00 240.1 F 10.2
26:00 247.3 F 14.0
28:00 258.8 F 20.4
30:00 269.2 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.6 F 29.4
2:00 271.9 F 28.6
3:00 271.7 F 29.1
4:00 272.8 F 29.3

Min 271.2 F 28.4
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.2 F 0.0
5:00 176.6 F 0.0
10:00 169.6 F 0.0
15:00 162.5 F 0.0
20:00 155.0 F 0.0
25:00 153.9 F 0.0
30:00 149.3 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:10
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:09
CYCLE COMPLETE

05/05/2022 08:14 AM
ACCEPTED: X
REJECTED:
Operator: U02