

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1700

Name: RITTER 1  
Sterilizer ID: V2178198  
03/09/2022 10:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 208.0 F 0.0  
2:00 199.3 F 3.3  
4:00 197.6 F 7.9  
6:00 224.1 F 11.1  
8:00 230.3 F 9.0  
10:00 238.2 F 9.7  
12:00 242.6 F 12.4  
14:00 241.6 F 11.1  
16:00 255.0 F 18.0  
18:00 266.6 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.0 F 28.6  
2:00 272.4 F 29.0  
3:00 271.7 F 28.8  
4:00 272.0 F 28.7  
  
Min 271.4 F 28.4  
Max 272.4 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.2 F 0.0  
5:00 232.8 F 0.0  
10:00 224.6 F 0.0  
15:00 226.3 F 0.0  
20:00 227.7 F 0.0  
25:00 231.1 F 0.0  
30:00 231.1 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 19:26  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:58:33  
CYCLE COMPLETE

03/09/2022 11:50 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_