
Midmark M11 - v1.0.0
Total Cycles: 3452

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2023 11:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.2 F 0.0
2:00 199.2 F 3.3
4:00 213.0 F 7.8
6:00 221.2 F 13.3
8:00 229.2 F 8.2
10:00 241.1 F 13.3
12:00 240.1 F 11.2
14:00 238.5 F 9.8
16:00 252.4 F 16.7
18:00 264.1 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.1 F 28.9
2:00 272.5 F 28.9
3:00 271.6 F 28.3
4:00 272.0 F 28.9

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 222.8 F 0.0
10:00 227.9 F 0.0
15:00 227.1 F 0.0
20:00 225.4 F 0.0
25:00 222.9 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 00:59:10
CYCLE COMPLETE

06/27/2023 12:09 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____