
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002818
15:46 09/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:35
STERILIZING 3:31
136.4 C 223kPa 3:31
136.2 C 224kPa 4:01
136.2 C 224kPa 4:31
136.3 C 221kPa 5:01
136.4 C 225kPa 5:31
136.3 C 223kPa 6:01
136.2 C 222kPa 6:31
136.4 C 221kPa 7:01
136.4 C 221kPa 7:01
VENTING 7:01
AIR DRYING 7:55
CYCLE COMPLETE 8:02

Digital Signature #
9601AB5CCBE7A2C3
