
Midmark M11 - v1.0.0
Total Cycles: 4578

Name: RITTER 1
Sterilizer ID: V2178198
05/24/2024 08:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.9 F 0.0
2:00 99.5 F 0.2
4:00 89.4 F 1.2
6:00 117.6 F 3.5
8:00 151.5 F 6.2
10:00 170.3 F 9.3
12:00 199.5 F 13.3
14:00 207.1 F 4.3
16:00 218.9 F 7.6
18:00 227.7 F 11.1
20:00 226.3 F 6.3
22:00 238.6 F 10.6
24:00 234.8 F 8.0
26:00 246.4 F 13.5
28:00 257.4 F 19.3
30:00 267.3 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.7 F 28.9
2:00 271.6 F 28.6
3:00 272.2 F 28.8
4:00 271.6 F 28.6

Min 271.4 F 28.3
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.7 F 0.0
5:00 231.1 F 0.0
10:00 226.9 F 0.0
15:00 206.4 F 0.0
20:00 211.5 F 0.0
25:00 213.0 F 0.0
30:00 207.9 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:30
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:09:38
CYCLE COMPLETE

05/24/2024 09:10 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____