
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006431
13:18 10/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:06
STERILIZING 4:05
136.0 C 220kPa 4:05
136.4 C 223kPa 4:35
136.4 C 225kPa 5:05
136.4 C 221kPa 5:35
136.0 C 219kPa 6:05
136.1 C 221kPa 6:35
136.4 C 224kPa 7:05
136.4 C 223kPa 7:35
136.4 C 223kPa 7:35
VENTING 7:35
AIR DRYING 8:30
CYCLE COMPLETE 10:34

Digital Signature #
E8261F42B4DAB74C
