
Midmark M11 - v1.0.5
Total Cycles: 2221

Name: RITTER2
Sterilizer ID: V2313228
04/10/2023 10:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.7 F 0.0
2:00 178.7 F 3.3
4:00 188.5 F 7.8
6:00 212.3 F 13.3
8:00 226.0 F 8.4
10:00 235.3 F 11.2
12:00 240.6 F 11.9
14:00 240.2 F 10.9
16:00 254.0 F 17.9
18:00 266.1 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.4 F 29.5
2:00 271.2 F 28.8
3:00 272.5 F 29.6
4:00 271.8 F 29.1

Min 271.1 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 223.4 F 0.0
10:00 224.9 F 0.0
15:00 227.1 F 0.0
20:00 230.1 F 0.0
25:00 230.4 F 0.0
30:00 229.7 F 0.0

Drying Complete

FILLING: 1:41
HEATING: 19:36
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

04/10/2023 11:06 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____