

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1118

Name: RITTER 1  
Sterilizer ID: V2178198  
08/25/2021 10:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.9 F 0.0  
2:00 188.4 F 4.3  
4:00 209.3 F 9.5  
6:00 222.2 F 6.6  
8:00 236.4 F 10.0  
10:00 235.1 F 8.2  
12:00 246.0 F 13.1  
14:00 244.8 F 12.5  
16:00 258.0 F 19.5  
18:00 269.6 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.3  
1:00 272.1 F 28.2  
2:00 272.4 F 28.6  
3:00 272.5 F 28.9  
4:00 271.8 F 28.2  
  
Min 271.7 F 28.1  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.4 F 0.0  
5:00 230.1 F 0.0  
10:00 218.8 F 0.0  
15:00 195.1 F 0.0  
20:00 194.3 F 0.0  
25:00 191.3 F 0.0  
30:00 189.9 F 0.0  
Drying Complete

FILLING: 2:12  
HEATING: 18:53  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:58:09  
CYCLE COMPLETE

08/25/2021 11:52 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02