

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1307

Name: RITTER2  
Sterilizer ID: V2313228  
07/15/2022 01:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.2 F 0.0  
2:00 169.2 F 0.1  
4:00 181.1 F 2.4  
6:00 205.7 F 7.0  
8:00 220.9 F 12.2  
10:00 227.3 F 8.0  
12:00 239.2 F 13.2  
14:00 239.4 F 11.2  
16:00 238.7 F 10.0  
18:00 251.7 F 16.6  
20:00 262.9 F 23.2  
22:00 272.6 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.0  
1:00 272.3 F 29.7  
2:00 271.7 F 28.9  
3:00 272.4 F 29.7  
4:00 271.9 F 29.0

Min 271.5 F 28.8  
Max 272.5 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.0 F 0.0  
5:00 226.2 F 0.0  
10:00 232.0 F 0.0  
15:00 226.7 F 0.0  
20:00 222.1 F 0.0  
25:00 220.2 F 0.0  
30:00 219.5 F 0.0  
Drying Complete

FILLING: 1:28  
HEATING: 22:07  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:44  
CYCLE COMPLETE

07/15/2022 02:36 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_