
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007174
10:15 09/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:04
STERILIZING 4:13
136.4 C 222kPa 4:13
136.2 C 224kPa 4:43
136.3 C 220kPa 5:13
136.4 C 220kPa 5:43
136.4 C 222kPa 6:13
136.4 C 225kPa 6:43
136.4 C 223kPa 7:13
136.3 C 224kPa 7:43
136.3 C 224kPa 7:43
VENTING 7:43
AIR DRYING 8:36
CYCLE COMPLETE 8:46

Digital Signature #
08F6D5388560E828
