

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 003459  
9:04 18/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:07  
PRESSURIZING 2:11  
STERILIZING 3:12  
136.0 C 220kPa 3:12  
135.8 C 221kPa 3:42  
136.4 C 221kPa 4:12  
136.3 C 222kPa 4:42  
136.4 C 221kPa 5:12  
136.4 C 224kPa 5:42  
136.3 C 223kPa 6:12  
136.4 C 225kPa 6:42  
136.4 C 225kPa 6:42  
VENTING 6:42  
AIR DRYING 7:36  
CYCLE COMPLETE 8:01  
-----

Digital Signature #  
BA4EAE1B531058AB  
-----