
Midmark M11 - v1.0.5
Total Cycles: 0

Name: RITTER2
Sterilizer ID: V2313228
04/21/2021 06:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.2 F 0.0
2:00 73.9 F 0.3
4:00 80.0 F 1.2
6:00 114.4 F 3.7
8:00 150.6 F 7.4
10:00 174.2 F 11.7
12:00 203.1 F 3.5
14:00 217.0 F 7.3
16:00 228.7 F 11.4
18:00 229.9 F 7.4
20:00 242.7 F 12.9
22:00 241.3 F 11.3
24:00 254.2 F 17.8
26:00 265.5 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 272.5 F 29.5
2:00 272.4 F 28.9
3:00 272.1 F 29.2
4:00 272.8 F 29.4

Min 271.8 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 165.9 F 0.0
Dry Cycle Aborted

FILLING: 0:53
HEATING: 27:35
STERILIZING: 4:00
VENTING: 2:51
DRYING: 2:30
TOTAL CYCLE: 00:37:50

04/21/2021 07:01 AM
ACCEPTED: X
REJECTED:
Operator: U01