

-----  
Midmark M11 - v1.0.0  
Total Cycles: 991

Name: RITTER 1  
Sterilizer ID: V2178198  
07/12/2021 06:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.7 F 0.0  
2:00 75.9 F 0.0  
4:00 77.0 F 0.0  
6:00 95.7 F 0.1  
8:00 130.5 F 0.1  
10:00 168.6 F 0.2  
12:00 192.0 F 1.6  
14:00 223.4 F 4.7  
16:00 217.6 F 7.6  
18:00 225.2 F 11.0  
20:00 232.6 F 7.3  
22:00 238.2 F 11.5  
24:00 236.6 F 8.7  
26:00 241.2 F 10.6  
28:00 246.7 F 13.2  
30:00 258.2 F 19.3  
32:00 268.4 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.2  
1:00 271.9 F 28.1  
2:00 272.2 F 28.2  
3:00 272.5 F 28.4  
4:00 272.7 F 28.7  
  
Min 271.7 F 28.1  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.6 F 0.0  
5:00 243.3 F 0.0  
10:00 239.7 F 0.0  
15:00 228.0 F 0.0  
20:00 223.7 F 0.0  
25:00 234.4 F 0.0  
30:00 225.6 F 0.0  
  
Drying Complete

FILLING: 0:44  
HEATING: 33:52  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:11:47  
CYCLE COMPLETE

07/12/2021 07:57 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01