
Midmark M11 - v1.0.5
Total Cycles: 1057

Name: RITTER2
Sterilizer ID: V2313228
05/02/2022 02:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.2 F 0.0
2:00 169.7 F 3.8
4:00 183.8 F 7.9
6:00 206.2 F 13.3
8:00 227.4 F 8.5
10:00 236.4 F 11.1
12:00 240.9 F 11.9
14:00 239.6 F 10.7
16:00 253.2 F 17.4
18:00 264.5 F 24.3
20:00 270.8 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.7
1:00 271.9 F 30.1
2:00 272.1 F 29.9
3:00 272.9 F 29.8
4:00 272.4 F 29.7

Min 271.6 F 29.6
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.1 F 0.0
5:00 239.8 F 0.0
10:00 230.6 F 0.0
15:00 227.5 F 0.0
20:00 229.0 F 0.0
25:00 226.6 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 0:36
HEATING: 20:32
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

05/02/2022 03:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____