
Midmark M11 - v1.0.5
Total Cycles: 897

Name: RITTER2
Sterilizer ID: V2313228
03/23/2022 12:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.7 F 0.0
2:00 157.1 F 0.1
4:00 185.5 F 1.5
6:00 206.6 F 6.5
8:00 221.2 F 11.7
10:00 229.6 F 8.0
12:00 242.4 F 12.5
14:00 242.6 F 12.4
16:00 242.7 F 11.9
18:00 256.6 F 19.2
20:00 268.7 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.2 F 29.2
2:00 271.1 F 28.9
3:00 272.3 F 29.2
4:00 271.0 F 28.7

Min 271.0 F 28.6
Max 272.4 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 228.3 F 0.0
10:00 218.3 F 0.0
15:00 196.5 F 0.0
20:00 192.9 F 0.0
25:00 215.0 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:05
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

03/23/2022 01:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____