
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003102
16:14 07/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:09
STERILIZING 4:08
136.4 C 221kPa 4:08
136.4 C 225kPa 4:38
136.4 C 225kPa 5:08
136.1 C 219kPa 5:38
136.2 C 221kPa 6:08
136.4 C 225kPa 6:38
136.4 C 225kPa 7:08
136.4 C 225kPa 7:38
136.4 C 225kPa 7:38
VENTING 7:38
AIR DRYING 8:33
CYCLE COMPLETE 8:39

Digital Signature #
3FA11CFD1AD7C4DB
