

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2409

Name: RITTER 1  
Sterilizer ID: V2178198  
10/06/2022 10:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.1 F 0.0  
2:00 177.7 F 2.5  
4:00 185.4 F 6.4  
6:00 218.9 F 12.6  
8:00 228.4 F 7.0  
10:00 237.7 F 12.4  
12:00 237.1 F 9.5  
14:00 237.2 F 8.9  
16:00 248.3 F 14.4  
18:00 260.5 F 21.1  
20:00 271.3 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.5 F 29.0  
2:00 271.4 F 28.2  
3:00 272.7 F 29.1  
4:00 272.0 F 28.5  
  
Min 271.3 F 28.2  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.8 F 0.0  
5:00 238.1 F 0.0  
10:00 224.2 F 0.0  
15:00 228.8 F 0.0  
20:00 231.1 F 0.0  
25:00 230.7 F 0.0  
30:00 231.0 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 20:40  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:26  
CYCLE COMPLETE

10/06/2022 11:26 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_